

## CLINICAL "SNIPPETS"

### **Improved chest recoil using an adhesive glove device for active compression-decompression CPR in a pediatric manikin model.**

Udassi JP, Udassi S, Lamb MA, Lamb KE, Theriaque DW, Shuster JJ, Zaritsky AL, Haque IU.  
Division of Pediatric Critical Care Medicine, Department of Pediatrics, University of Florida College of Medicine, Gainesville, FL, United States; Division of Pediatric Cardiology, Department of Pediatrics, University of Florida College of Medicine, Gainesville, FL, United States.

*Resuscitation. 2009 Aug 14.*

**Overview:** We developed an adhesive glove device (AGD) to perform ACD-CPR in pediatric manikins, hypothesizing that AGD-ACD-CPR provides better chest decompression compared to standard (S)-CPR.

**Conclusions:** A simple, inexpensive glove device for ACD-CPR improved chest decompression with emphasis on active pull in manikins without excessive rescuer fatigue. The clinical implication of fewer compressions/minute in the AGD group needs to be evaluated.